

Paragon²⁰

a  ZIMMER BIOMET company



SURGICAL TECHNIQUE GUIDE

Tibiotalar/Tibiotalocalcaneal Arthrodesis



SILVERBACKTM
ANKLE FUSION PLATING SYSTEM

PRODUCT DESCRIPTION

The Paragon 28® Silverback™ Ankle Fusion Plating System was designed to give surgeons options for tibiotalar (TT) and tibiotalocalcaneal (TTC) arthrodesis. The system includes anterior, lateral, and posterior approach specific plating families helps accommodate variations in anatomy for both TT and TTC versions.

Both the Anterior and Posterior plates include two forms, a flat and contoured version to help accommodate variations in anatomy for both TT and TTC versions. The Lateral plates are available in TT and TTC versions. To accommodate arthroscopic or mini-open fusion procedures, an Anterior Mini-Open Plate is available to be used in combination with cannulated screws for increased construct rigidity and decreased micromotion at the tibiotalar joint when compared to cannulated screws in isolation. The relatively thinner plates help to evenly distribute force across the construct and guard against stress shielding during healing.

Plate hole sizing allows for Ø4.5 mm and Ø5.2 mm screws to be used for the tibia and calcaneus, while the talar screw holes allow for Gorilla® R3CON Ø3.5 mm and Ø4.2 mm screws, with the exception of the Mini-Open Plate which uses a Silverback screw in the talar body. A Ø4.7 mm “Compact” screw is available for the hole sizes in the tibia and calcaneus, which was designed with a smaller thread height to help reduce insertion torque in dense bone. Additionally, single lead bone threads result in a decreased pitch differential between the locking screw head and bone threads to reduce the amount of insertion torque required to lock the screw into the plate in areas of dense bone. Precision® Guides are provided to allow for a crossing screw to be inserted outside of the plate while avoiding interference with the on-axis plate screws. Each plate has one Precision® Guide to place a tibiotalar screw, while the Lateral and Posterior TTC plates have an additional Precision® Guide to place the subtalar crossing screw. The Anterior Mini-Open Plates have two Precision® Guides to provide placement of independent medial and lateral crossing screws across the tibiotalar joint.


TABLE OF CONTENTS

INTRODUCTION	IMPLANTS: PLATE FAMILIES	3
	FEATURED INSTRUMENTS	4-6
	SLIM LINE SYSTEM	7-8
FEATURED TECHNIQUES	ANTERIOR TIBIOTALAR ARTHRODESIS.....	9-13
	ANTERIOR TIBIOTALOCALCANEAL ARTHRODESIS	14-17
	ANTERIOR MINI-OPEN TIBIOTALAR ARTHRODESIS	18-24
	LATERAL TIBIOTALOCALCANEAL ARTHRODESIS	25-28
	LATERAL TIBIOTALAR ARTHRODESIS	29
	POSTERIOR TIBIOTALAR ARTHRODESIS	30-34
	POSTERIOR TIBIOTALOCALCANEAL ARTHRODESIS	35-40
APPENDIX	CADDY LAYOUT	41-47
	MR SAFETY AND INFORMATION.....	48-49

ANKLE FUSION PLATES


Available in Right (shown) and Left Configurations

Anterior TT Plates



Anterior TTC Plates

Central talar body hole allows for a Ø7.0 mm Monster® Screw across the subtalar joint




Side Profile



Flat Anterior Plate

Contoured Anterior Plate

Anterior Mini-Open Plate



Side Profile



Flat Anterior Mini-Open Plate

Contoured Anterior Mini-Open Plate

Lateral TT Plates



Lateral TTC Plates




Side Profile




Lateral TT Plate

Lateral TTC Plate


Posterior TT Plates



Posterior TTC Plates



Side Profile



Flat Posterior Plate

Contoured Posterior Plate

FEATURED INSTRUMENTS

	Ø3.5 mm R3CON Screws	Ø4.2 mm R3CON Screws	Ø4.5 mm SILVERBACK™ Screws	Ø5.2 mm SILVERBACK™ Screws	Ø4.7 mm SILVERBACK™ Compact Screws
Locking:					
Non-locking:					
Screw Lengths:	14 mm - 30 mm in 2 mm increments	14 mm - 50 mm in 2 mm increments and 55 mm - 60 mm in 5 mm increments			20 mm - 40 mm in 2 mm increments
Drill Size:	Ø2.4 mm 	Ø2.8 mm 	Ø3.1 mm 	Ø3.6 mm 	Ø3.6 mm 
Driver Size:	HX-10 	HX-10 	HX-15 	HX-15 	HX-15 
Locking Drill Guide Size	Ø3.5 mm 	Ø4.2 mm 	Ø4.5 mm 	Ø4.7/Ø5.2 mm 	Ø4.7/Ø5.2 mm 
Centering Drill Guide Size	Ø3.5 mm 	Ø4.2 mm 	Ø4.5 mm 	Ø5.2 mm 	N/A
Compression Slot Drill Guide Size:	N/A	N/A	Ø4.5 mm 	Ø4.7/Ø5.2 mm 	Ø4.7/Ø5.2 mm 
Cone/Straight Easy Guide Size:	Ø3.5 mm 	Ø4.2 mm 	Ø4.5 mm 	Ø4.7/Ø5.2 mm 	Ø4.7/Ø5.2 mm 

Drill



Driver



Locking Drill Guide



Centering Drill Guide



Compression Slot Drill Guide



Cone/Straight Easy Guide



JOINT PREPARATION INSTRUMENTATION

Bone Fenestration Perforator



Curved Bone Fenestration Chisel



Cartilage Removal Tool



Angled Ring Curette



Straight Ring Curette



Angled Curette



Straight Curette



Oval Burr



Barrel Burr



Straight Bone Fenestration Chisel



Curved 3 mm Osteotome



Straight 6 mm Osteotome



Curved 6 mm Osteotome



Straight 12 mm Osteotome



Curved 12 mm Osteotome



OTHER INSTRUMENTATION

Ø2.5 & Ø3.0 mm K-wire Hindfoot Distractor



Overdrill Guide




AO Handles



Threaded Plate Bending Bars



Flat Plate Benders



Ø2.0 x 200 mm K-wire (Smooth and Threaded)



Ø2.5 x 150 mm K-wire (Smooth and Threaded)



Olive Wire, Short (Smooth and Threaded)




Olive Wire, Long (Smooth and Threaded)




JOINT PREPARATION INSTRUMENTATION


Anterior TT/TTC Precision Guide




Anterior Mini-Open Instrumentation



Lateral TT/TTC Precision Guide



Posterior Precision Guides




SILVERBACK SLIMLINE SYSTEM

The Silverback Ankle Fusion Plating System is available in a condensed Slim Line version case, containing a curated selection of instruments and implants for use in the most routinely encountered Silverback plating procedures. The Silverback Slim Line implant and instrument contents are shown below.

PLATES


Available in Right (shown) and Left Configurations

Anterior TT Plates




Anterior TTC Plates

Central talar body hole allows for a Ø7.0 mm Monster® Screw across the subtalar joint



Anterior Mini-Open Plate



SCREWS

	Ø3.5 mm R3CON Screws	Ø4.2 mm R3CON Screws	Ø4.5 mm SILVERBACK™ Screws	Ø5.2 mm SILVERBACK™ Screws
Locking:				
Non-locking:				
Screw Lengths:	2 mm increments 14-30 mm	2 mm increments 14-40 mm	2 mm increments 14-50mm, 5 mm increments 55-60 mm	
Drill Size:	Ø2.4 mm	Ø2.8 mm	Ø3.1 mm	Ø3.6 mm
Driver Size:	HX-10	HX-10	HX-15	HX-15
Locking Drill Guide Size:	Ø3.5 mm	Ø4.2 mm	Ø4.5 mm	Ø4.7/Ø5.2 mm
Compression Slot Drill Guide Size:	N/A	N/A	Ø4.5 mm	Ø4.7/Ø5.2 mm
Cone/Straight Easy Guide Size:	Ø3.5 mm	Ø4.2 mm	Ø4.5 mm	Ø4.7/Ø5.2 mm

For each screw size
listed above

Drill

Driver

Locking Drill Guide

Centering Drill Guide


Compression Slot Drill Guide

Cone/Straight Easy Guide

SILVERBACK SLIM LINE SYSTEM

INSTRUMENTS

Flat Plate Benders



Threaded Plate Bending Bars



Bone Fenestration Perforator



Straight Bone Fenestration Chisel



Curved Bone Fenestration Chisel



Ø2.0 x 200 mm K-wire (Smooth)



Olive Wire, Short (Smooth and Threaded)



Cartilage Removal Tool



Angled Ring Curette





Angled Curette



Curved 12 mm Osteotome



Anterior TT/TTC Precision Guide

Subtalar Joint
Wire GuidePrecision Guide K-wire
Tube, Ø1.6 mmPrecision Guide K-wire
Tube, Ø2.3 mmPrecision Guide Arm,
Flat PlatesPrecision Guide Arm,
Contoured Plates


Anterior Mini-Open Instrumentation

Precision Guide
K-wire Tube, Ø1.6 mmPrecision Guide
K-wire Tube, Ø2.3 mmOlive Wire, Short Mini-Open
(Smooth and Threaded)Threaded Tower (Joystick) Olive Wire,
Long (Smooth and Threaded)

Contoured Precision Guide Base



Flat Precision Guide Base

Threaded Tower
(Joystick)Precision Guide Arm,
LateralPrecision Guide Arm,
Medial

INCISION/EXPOSURE

TIBIOTALAR JOINT

A longitudinal midline incision is made over the anterior ankle, beginning approximately 10 cm proximal to the ankle joint and terminating just distal to the talonavicular joint. The incision will start approximately 1 cm lateral to the tibial crest and will course just lateral to the tibialis anterior (TA) tendon. The initial incision should penetrate skin only, but no direct tension should be placed on the skin margins until full-thickness retraction is possible. Identify the superficial peroneal nerve and retract it laterally. Continue exposure to the extensor retinaculum. Identify the extensor hallucis longus (EHL) tendon below the retinaculum and divide the retinaculum longitudinally over the EHL tendon. Care should be taken to leave the sheath of the TA tendon intact and the retinaculum well preserved for repair at closure.



Retract the EHL tendon laterally and the TA tendon medially. Identify the neurovascular bundle and retract it laterally with the EHL tendon. Continue exposure until the anterior capsule is visualized. Perform an anterior capsulotomy via a longitudinal incision. Elevate the capsule and periosteum over the anterior tibia and talus to expose the anterior ankle joint, the tibial plafond, the medial and lateral gutters and the anterior and dorsal talus. Remove any osteophytes on the tibia and talus to allow for exposure to the ankle joint and facilitate entry of instrumentation for cartilage removal. All osteophytes must be removed from the anterior ankle to facilitate application of the plate.

JOINT PREPARATION

Preparation of the tibiotalar joint can be performed using the provided joint preparation instrumentation. A Hindfoot Distractor is offered to allow for space and visualization during joint preparation, to be used with the Ø2.5 mm K-wires. The Hindfoot Distractor and K-wires are placed through the lateral aspect of the incision with appropriate soft tissue retraction, in an anterior to posterior direction. Following cartilage removal, it is advised to penetrate the subchondral plate with the Subchondral Drill, Burrs and/or Bone Fenestration Chisels to promote healing.

PROVISIONAL FIXATION

Align the ankle joint. The foot and ankle should be positioned such that the ankle is neutral with respect to dorsiflexion and plantarflexion. The foot should be in approximately 5-10° of external rotation and 5° of hindfoot valgus. With the foot and ankle held in this alignment, use one or two Ø2.0 mm K-wires to temporarily fixate the tibiotalar joint, per surgeon preference.



PLATE PLACEMENT



Frequently there is an irregular surface remaining over the anterior joint which requires debridement to a smooth surface with either a rongeur or saw. Retrieve the appropriate anterior TT plate based on the patient's anatomy. To position the plate, palpate the medial and lateral margins of the talus and center the talar portion of the plate. Ensure that the proximal plate is midline or just lateral to midline.

Secure the plate to the tibiotalar joint using a long Olive Wire in the most proximal circular hole on the tibia and a short Olive Wire in the medial talar neck screw hole. Confirm plate position using fluoroscopy.

PERMANENT FIXATION - PLATE SCREWS



NOTE: The talar screw holes accept Ø3.5 mm or Ø4.2 mm non-locking and locking screws. Ø4.2 mm screws are recommended for this area, except in the case of a small patient. The use of Ø4.2 mm screws is demonstrated in this technique. When using Ø3.5 mm screws, use the appropriate instrumentation as described on page 4.



Retrieve the Ø4.2 mm Locking Drill Guide and thread into the lateral talar body screw hole. Drill, using the Ø2.8 mm Drill.



Remove the Ø4.2 mm Locking Drill Guide and measure screw length using the Depth Gauge. Confirm screw projection and length using the Depth Gauge under fluoroscopy (not shown). Insert the selected screw size into the plate hole using the provided Driver and Handle. Do not fully tighten screw until the second talar screw is secure, to prevent toggling of the plate.



Remove the Olive Wire in the talar neck hole. Insert a second Ø4.2 mm screw into the medial talar body hole using the same procedure previously described. Complete tightening and seating of both talar body screws.

PERMANENT FIXATION - TIBIAL COMPRESSION SCREW



NOTE: The tibial screw holes accept Ø4.5 mm, Ø4.7 mm, or Ø5.2 mm locking or non-locking screws. A laser etched dot on the plate indicates the plate holes that accept the Ø4.5 mm, Ø4.7 mm, and Ø5.2 mm screws. The technique demonstrates the use of the Ø4.5 mm screws. When using the Ø4.7 mm or Ø5.2 mm screws, use the appropriate instrumentation as described on page 4.



Retrieve the Ø4.5 mm oblong Compression Slot Drill Guide and insert into the tibial compression slot with the arrow pointing towards the tibiotalar joint. Drill, using a Ø3.1 mm Drill through the Compression Slot Drill Guide.




Remove the Compression Slot Drill Guide and measure screw length using the Depth Gauge. Insert a Ø4.5 mm non-locking screw. Do not fully tighten the screw until the tibiotalar crossing screw has been inserted. This allows for travel of the compression screw distally in the slot during final tightening of the crossing screw.

PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREW

Retrieve the corresponding anterior TT Precision Guide and Set Screw. Insert the provided Set Screw into the hole adjacent to the boss on the undersurface of the Precision Guide Arm.

Align the boss on the underside of the Precision Guide into the smaller of the two central holes on the ankle fusion plate to correctly orient the Precision Guide, while aligning the Set Screw over the larger of the two central holes. Rotate the Set Screw clockwise to secure the Precision Guide to the ankle fusion plate.



NOTE: Alternatively, the Precision Guide can be placed on the plate following completion of plate fixation to bone. Placement of a fully threaded crossing screw would take place after fully seating the tibial non-locking screw in the compression slot.

PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREW



NOTE: K-wire Tubes for the Precision Guide are available in Ø1.6 mm and Ø2.3 mm, allowing for Ø5.5 mm or Ø7.0 mm Monster Screws to be used. Partially threaded and fully threaded screw options are available for each screw diameter, per surgeon preference. The use of Ø7.0 mm Monster Screw is demonstrated in this technique. When using the Ø5.5 mm Monster Screw, use the corresponding instrumentation for that diameter screw.



Retrieve the Ø2.3 mm K-wire for a Ø7.0 mm partially threaded Monster Screw. Insert the K-wire through the K-wire Tube and across the arthrodesis site.

Confirm crossing screw trajectory using fluoroscopy. The ideal position for the trajectory is terminating just proximal to the lateral process of the talus.



Remove the K-wire sleeve from the Ø2.3 mm K-wire. Remove the Precision Guide by turning the Set Screw counterclockwise to detach the Precision Guide from the plate. Remove the Precision Guide and slide the K-wire Tube off the K-wire.



Retrieve the Countersink for the Ø7.0 mm headed Monster Screw. Rotate the Countersink clockwise over the K-wire to remove adequate bone to seat the screw head. Measure screw length using the Depth Gauge (not shown).



NOTE: If soft bone is apparent, it is advised not to countersink in this area to allow for better screw purchase. Monster Screw washers are available; however, the steep screw angle may cause prominence of the washer.




TIP: If minimal compression or poor bone purchase occurs with this crossing screw, remove the partially threaded screw and guide wire. Obtain compression by fully tightening the non-locking screw in the compression slot and then place a fully threaded screw via the Precision Guide.

PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREW



Drill over the K-wire using the Ø4.6 mm Drill for the Ø7.0 mm headed Monster Screw.



Insert a Ø7.0 mm headed Monster Screw using the provided Driver. Prior to seating the Monster Screw head against bone, remove the Olive Wire from the tibial screw hole and the provisional fixation across the tibiotalar joint. Fully seat the Monster Screw, then confirm screw length and placement using fluoroscopy. Remove the Ø2.3 mm K-wire.



Fully seat the non-locking screw in the tibia compression slot.

PERMANENT FIXATION - PLATE SCREWS



Insert two Ø4.2 mm screws into the talar neck holes using the same procedure previously described.



Insert the remaining tibial screws using the technique previously described for Ø4.5 mm screws. Confirm screw lengths and placement using fluoroscopy.



CLOSURE

Proceed to incision closure or concomitant procedures at this time.

INCISION/EXPOSURE

TIBIOTALAR JOINT

An incision for the tibiotalar joint is made as described on page 7 for the anterior TT arthrodesis surgical technique.

SUBTALAR JOINT

A longitudinal incision is made starting at the distal aspect of the lateral malleolus over the subtalar joint extending toward the 4th metatarsal base and terminating at the calcaneocuboid joint. Continue exposure through the subcutaneous tissue, with care being taken to identify and retract the anterior branch of the sural nerve. The extensor hallucis brevis muscle is reflected distally to expose the sinus tarsi and posterior facet of the subtalar joint. Dissection of the fat pad out of the sinus tarsi should be performed, with reflection of the tissue dorsally.



JOINT PREPARATION

Preparation of the tibiotalar joint can be performed using the provided joint preparation instrumentation. The medial and lateral gutters should be cleared at this time. Remove cartilage from the posterior, middle and anterior facets of the subtalar joint. A Hindfoot Distractor is provided to allow for space and visualization during joint preparation, to be used with provided Ø2.5 mm K-wires. It is advised to penetrate the subchondral plate with the Subchondral Drill, Burrs and/or Chisels to promote healing.

PROVISIONAL FIXATION

Align the ankle joint. The foot and ankle should be positioned such that the ankle is neutral with respect to dorsiflexion and plantarflexion. The foot should be in approximately 5-10° of external rotation and 5° of hindfoot valgus. With the foot and ankle held in this alignment, use Ø2.0 mm K-wires to temporarily fixate the tibiotalar joint and subtalar joint, per surgeon preference.




PLATE POSITIONING




Resection of osteophytes or prominent bone may be necessary to ensure proper plate fit. Retrieve the appropriate anterior TTC plate based on the patient's anatomy.

To position the plate, palpate the medial and lateral margins of the talus and center the talar portion of the plate. Ensure that the proximal plate is midline or just lateral to midline.

Secure the plate to the tibiotalar joint using a long Olive Wire in the most proximal circular hole on the tibia and a short Olive Wire in the medial talar neck screw hole. Confirm plate position using fluoroscopy.



PERMANENT FIXATION – TALAR NECK AND TIBIAL COMPRESSION SCREWS




Insert talar neck screws as described on page 8. Insert the tibial compression screw as described on page 9.

PERMANENT FIXATION – TALAR NECK AND TIBIAL SCREWS



Assemble and attach the Precision Guide to the anterior TTC plate as described on page 9.




Insert the tibiotalar crossing screw as described on page 10 and 11. Prior to seating the Monster Screw head against bone, remove the Olive Wire from the tibial screw hole and the provisional fixation across the tibiotalar joint. Fully seat the Monster Screw, then confirm screw length and placement using fluoroscopy. Remove the Ø2.3 mm K-wire.

PERMANENT FIXATION – TIBIAL SCREWS




Insert tibial screws as described on page 11.

PERMANENT FIXATION – SUBTALAR JOINT CROSSING SCREW




A Ø7.0 headed Monster Screw is used in this plate hole. Retrieve the Subtalar Joint Wire Guide for a Ø7.0 mm headed Monster Screw. Mate the tip of the Wire Guide into the socket of the plate.



Insert a Ø2.3 mm K-wire through the Wire Guide across the subtalar joint, aiming for the central posterior aspect of the calcaneus, just above the weight bearing surface.

Remove the K-wire serving as provisional fixation across the subtalar joint.

Measure screw length using the cannulated Depth Gauge.



Drive the Ø2.3 mm K-wire posteriorly until it exits the skin. Retrieve a Parallel K-wire Guide from the Monster Screw System instrument caddy. Slide the central, isolated hole of the Parallel K-wire Guide over the Ø2.3 mm K-wire.




Depending on screw type used (headed vs. headless), place a second Ø2.3 mm K-wire a desired distance proximal to the initial Ø2.3 mm K-wire.


PERMANENT FIXATION – SUBTALAR JOINT CROSSING SCREW



NOTE: A partially threaded screw is recommended to be placed in the more proximal position, to allow for the threads of the screw to engage the denser bone in the talus and create compression across the subtalar joint. The second screw placed in the plate is intended to be a fully threaded screw to hold in the compression created from the first screw.




Placement of a partially threaded Ø7.0 mm Monster Screw through the posterior calcaneus into the talus is performed as described on pages 10 and 11. Remove the Ø2.3 mm K-wire. Confirm screw placement using fluoroscopy.



NOTE: Countersinking is not necessary, as the Ø7.0 mm Monster Screw is seated within the plate.

Drill over the K-wire using the Ø4.6 mm Drill for the Ø7.0 mm headed Monster Screw.




Insert a Ø7.0 mm headed, fully threaded Monster Screw using the provided Driver.

Confirm screw length and placement using fluoroscopy. Remove the Ø2.3 mm K-wire.

CLOSURE

Proceed to incision closure or concomitant procedures at this time.



INCISION/EXPOSURE

A longitudinal midline incision is made over the anterior ankle, beginning approximately 2 cm proximal to the tibiotalar joint, and terminating just distal to the talonavicular joint. The incision will start approximately 1 cm lateral to the tibial crest and will course just lateral to the tibialis anterior (TA) tendon. The initial incision should penetrate skin only, but no direct tension should be placed on the skin margins until full-thickness retraction is possible. Identify the superficial peroneal nerve and retract it laterally. Continue exposure to the extensor retinaculum. Identify the extensor hallucis longus (EHL) tendon below the retinaculum and divide the retinaculum longitudinally over the EHL tendon. Care should be taken to leave the sheath of the TA tendon intact and the retinaculum well preserved for repair at closure.

Retract the EHL tendon laterally and the TA tendon medially. Identify the neurovascular bundle and retract it laterally with the EHL tendon. Continue exposure until the anterior capsule is visualized. Perform an anterior capsulotomy via a longitudinal incision. Elevate the capsule and periosteum over the anterior tibia and talus to expose the anterior ankle joint, the tibial plafond, the medial and lateral gutters and the anterior and dorsal talus. Remove any osteophytes on the tibia and talus to allow for exposure to the ankle joint and facilitate entry of instrumentation for cartilage removal. All osteophytes must be removed from the anterior ankle to facilitate application of the plate.



JOINT PREPARATION

Preparation of the tibiotalar joint can be performed using the provided joint preparation instrumentation. A Hindfoot Distractor is offered to allow for space and visualization during joint preparation, to be used with the Ø2.5 mm K-wires. The Hindfoot Distractor and K-wires are placed through the lateral aspect of the incision with appropriate soft tissue retraction, in an anterior to posterior direction. Following cartilage removal, it is advised to penetrate the subchondral plate with the Subchondral Drill, Burrs and/or Bone Fenestration Chisels to promote healing. Alternatively, the ankle joint may be prepared using arthroscopic instrumentation per surgeon preference.

PROVISIONAL FIXATION


Align the ankle joint. The foot and ankle should be positioned such that the ankle is neutral with respect to dorsiflexion and plantarflexion. The foot should be in approximately 5-10° of external rotation and 5° of hindfoot valgus. With the foot and ankle held in this alignment, place a 2.0 mm K-wire lateral to medial for temporary fixation the tibiotalar joint.



PLATE POSITIONING AND PROVISIONAL FIXATION



Prior to the temporary fixation of the plate, thread the Threaded Tower (Joystick) through the medial tibial screw hole (as shown) to aid in plate positioning. To position the plate, palpate the medial and lateral margins of the talus and center the talar portion of the plate. Ensure that the proximal plate is midline or just lateral to midline.



Secure the plate to the tibiotalar joint by placing a Long Olive Wire from the Anterior Mini-Open Plate Caddy through the Threaded Tower (Joystick) in the medial tibial screw hole and an Olive Wire from the Silverback K-wire and Olive Wire Caddy in the medial talar neck screw hole. Confirm plate position using fluoroscopy.


PERMANENT FIXATION – PLATE SCREWS



NOTE: The talar body screw holes accept Ø4.5 mm Ø4.7 mm, or Ø5.2 mm locking or non-locking screws. A laser etched dot on the plate indicates the plate holes that accept Ø4.5 mm, Ø4.7 mm, and Ø5.2 mm screws. The technique demonstrates the use of the Ø4.5 mm screws. When using the Ø4.7 mm or Ø5.2 mm screws, use the appropriate instrumentation as described on page 4. The appropriate Locking Drill Guide is recommended for all tibia screws and the talar body screw to achieve on-axis trajectory and avoid interference with the crossing screws.



Retrieve the Ø4.5 mm Threaded Drill Guide and thread into the talar body screw hole. Drill, using the Ø3.1 mm Drill.



Remove the Ø4.5 mm Threaded Drill Guide and measure screw length using the Depth Gauge (not shown). Confirm screw projection and length using the Depth Gauge under fluoroscopy. Insert and fully seat the selected talar body screw into the plate hole using the provided Driver and Handle.




Remove the Olive Wire in the talar neck hole. Do not remove the Olive Wire from the Threaded Tower (Joystick) of the tibia to prevent shifting of the plate.


PERMANENT FIXATION – TIBIAL COMPRESSION SCREW



NOTE: The tibial screw holes accept Ø4.5 mm Ø4.7 mm, or Ø5.2 mm locking or non-locking screws. A laser etched dot on the plate indicates the plate holes that accept Ø4.5 mm, Ø4.7 mm, and Ø5.2 mm screws. The technique demonstrates the use of the Ø4.5 mm screws. When using the Ø4.7 mm or Ø5.2 mm screws, use the appropriate instrumentation as described on page 4. The appropriate Locking Drill Guide is recommended for all tibia screws and the talar body screw to achieve on-axis trajectory and avoid interference with the crossing screws.




Retrieve the Ø4.5 mm oblong Compression Slot Drill Guide and insert into the tibial compression slot with the arrow pointing towards the tibiotalar joint. Drill, using a Ø3.1 mm Drill through the Compression Slot Drill Guide.




Remove the Compression Slot Drill Guide and measure screw length using the Depth Gauge. Insert a Ø4.5 mm non-locking screw. Do not fully tighten the screw until the medial tibiotalar crossing screw has been inserted. This allows for travel of the compression screw distally in the slot during final tightening of the crossing screw. Remove the Long Olive Wire from the tibia to allow for removal of the Threaded Tower (Joystick) at this time.

PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREWS

Prior to the placement of the Precision Guide Base on to the plate, attach the Medial Precision Guide Arm to the Precision Guide Arm Base. Align the pin of the Medial Precision Guide Arm with the proximal hole located on the medial side of the Precision Guide Base. Secure the Guide Arm to the Arm Base by turning the set screw clockwise.





NOTE: The Medial Precision Guide Arm is labeled using laser etching and has a small notch on the top ridge of the guide. This technique demonstrates the use of a right foot.



The Precision Guide Base is universal for right and left sides and is symmetric about its mid-line. The Medial and Lateral Precision Guide Arms will drive the trajectory of the guide wire for the crossing screw.

PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREWS

After connecting the Medial Precision Guide Arm to the Precision Guide Base, the Arm Base inserts into the plate through two holes. Align the boss on the underside of the Precision Guide Base with the tibia screw hole located distal to the compression slot. Align the distal through-hole of the Base with the small threaded hole located in the center of the plate. A Set Screw will be used to secure the base to the plate by rotating the Set Screw clockwise.





NOTE: The Medial Precision Guide Arm allows for variation anteriorly and posteriorly in the placement of the crossing screw, depending on patient size and anatomy. It is recommended to use the more posterior crossing screw trajectory, if the patient anatomy allows, for a more balanced construct.



NOTE: K-wire Sleeves for the Precision Guides are available in Ø1.6 mm and Ø2.3 mm, allowing for Ø5.5 mm or Ø7.0 mm Monster crossing screws to be used. Partially threaded and fully threaded screw options are available for each screw diameter, per surgeon preference. The use of Ø7.0 mm Monster crossing screws are demonstrated in this technique. When using the Ø5.5 mm Monster Screw, use the corresponding instrumentation for that diameter screw.

Place the K-wire Sleeve into the Medial Precision Guide Arm to allow placement of the Ø2.3 mm K-wire, serving as a guide wire for the Ø7.0 mm Monster Screw.

Place a Ø2.3 mm K-wire through the K-wire Sleeve and insert across the tibiotalar joint.



Confirm K-wire placement and trajectory using fluoroscopy.

Remove the K-wire Sleeve from the Ø2.3 mm K-wire. Remove the Precision Guide Arm from the Precision Guide Base, but do not remove the Precision Guide Base at this time.


PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREWS



Retrieve the Countersink for the Ø7.0 mm partially threaded Monster Screw. Rotate the Countersink clockwise over the K-wire to remove adequate bone to seat the screw head. Measure screw length using the Depth Gauge (not shown).




Drill over the K-wire using the Ø4.6 mm Drill for the Ø7.0 mm partially threaded Monster Screw.



Insert, but do not fully seat the Ø7.0 mm partially threaded Monster Screw using the provided Driver. Remove the Ø2.0 mm K-wire serving as provisional fixation from the tibia before fully seating the Monster Screw.



Fully seat the Monster Screw. Confirm screw length and placement using fluoroscopy.




Once the Monster Screw is fully seated, use the driver for the Ø4.5 mm screws to fully seat the non-locking screw in the compression slot of the tibia.


Attach the Lateral Precision Guide Arm to the Precision Guide Base in the same manner as described for the Medial Precision Guide Arm. Repeat the process of K-wire placement that was used for the medial Monster Screw.



NOTE: The Lateral Precision Guide Arm allows for variation in anterior and posterior trajectories that converge at the medial aspect of the talus. The trajectories are biased anteriorly at the entry in the tibia to help avoid the fibula. It is recommended to use the more posterior crossing screw trajectory, if the patient anatomy allows, for a more balanced construct



PERMANENT FIXATION - PRECISION GUIDED CROSSING SCREWS



After the K-wire has been placed across the tibiotalar joint, remove the K-wire Sleeve. The Precision Guide Arm Base and the Lateral Precision Guide Arm may be removed simultaneously by turning the Precision Guide Base Set Screw counterclockwise to detach from the Mini Open plate.




Using the previous technique described on pages 19-20 for the medial partially threaded Monster Screw, place a Ø7.0 mm fully threaded Monster Screw. Remove the Ø2.3 mm K-wire.


PERMANENT FIXATION – PLATE SCREWS

NOTE: The talar neck screw holes accept Ø3.5 mm or Ø4.2 mm non-locking and locking screws. The use of Ø4.2 mm screws is demonstrated in this technique. When using the Ø3.5 mm screws, use the appropriate instrumentation as described on page 4.


Retrieve the Ø4.2 mm Locking Drill Guide and thread into the lateral talar neck screw hole. Drill, using the Ø2.8 mm Drill. Remove the Ø4.2 mm Locking Drill Guide and measure screw length using the Depth Gauge. Confirm screw projection and length using the Depth Gauge under fluoroscopy (not shown).



PERMANENT FIXATION – PLATE SCREWS



Insert the selected screw in the plate using the provided Driver and Handle. Insert the other remaining talar neck screw using the technique previously described for Ø4.2 mm screws. Confirm screw lengths and placement using fluoroscopy.




Retrieve the Ø4.5 mm Locking Drill Guide and thread into the tibial screw hole. Drill, using the Ø3.1 mm Drill. Remove the Ø4.5 mm Locking Drill Guide and measure screw length using the Depth Gauge.



Insert the remaining tibial screws using the technique previously described for Ø4.5 mm screws. Confirm screw projection and length under fluoroscopy (not shown).

CLOSURE

Proceed to incision closure or concomitant procedures at this time.



INCISION/EXPOSURE

INCISION/EXPOSURE

A lateral incision is made over the posterior half of the fibula, beginning approximately 10 cm proximal to the tip of the fibula extending distally to the plantar aspect of the calcaneus.

Identify the sural nerve and retract it posteriorly. Continue dissection to the fibula. While retracting the peroneal tendons and sural nerve, a transverse fibular osteotomy is performed by beveling the saw from proximal lateral to distal medial to avoid a sharp bony prominence above the plate. Transect the syndesmotic and lateral ankle ligaments to free the fibula from adjacent soft tissues. Resect the fibula and retain for bone graft, if desired. Alternatively, if available, a reamer can be used to remove the fibula, while harvesting the reamed bone for graft material.

Elevate the anterior joint capsule and nearby periosteum to assess the anterior tibiotalar joint articulation. Minimal dissection of the talar neck is recommended to avoid devascularization of this bone. Remove any anterior osteophytes that may interfere with joint reduction. Elevate the posterior soft tissues using a periosteal elevator to allow for retractors to be placed anterior to and posterior to the tibiotalar joint. Elevate the extensor digitorum brevis muscle belly to expose the subtalar joint. Release the lateral ligaments around the subtalar joint including the talocalcaneal intraosseous ligament to allow for appropriate distraction. Dissect the fat pad out of the sinus tarsi to allow for appropriate visualization.



JOINT PREPARATION

Preparation of the tibiotalar joint can be performed using the provided joint preparation instrumentation. A Hindfoot Distractor is offered to allow for space and visualization during joint preparation, to be used with the Ø2.5 mm K-wires. It is advised to penetrate the subchondral plate with the Subchondral Drill, Burrs and/or Chisels to promote healing. A medial arthrotomy may be required and performed to allow for exposure and joint preparation of the medial gutter of the ankle joint.

Remove cartilage from the posterior, middle and anterior facets of the subtalar joint. Perform subchondral plate penetration to these joints to promote healing.

PROVISIONAL FIXATION

Align the ankle joint. The foot and ankle should be positioned such that the ankle is neutral with respect to dorsiflexion and plantarflexion. The foot should be in approximately 5-10° of external rotation and 5° of hindfoot valgus. With the foot and ankle held in this alignment, use multiple Ø2.0 mm K-wires to temporarily fix the tibiotalar and subtalar joints as shown. Provisional fixation wires should be placed in the anterolateral to posteromedial direction for the tibiotalar joint, and posterior to anterior direction for the subtalar joint.




PLATE POSITIONING AND PROVISIONAL FIXATION



Resection of osteophytes or prominent bone such as the lateral talar process may be necessary to ensure proper plate fit. Retrieve the appropriate lateral TTC plate based on the patient's anatomy. The plate should be positioned such that the proximal aspect is centered from anterior to posterior on the tibia, the central talar hole is centered on the body of the talus and the posterior calcaneal holes align just inferior to the superior surface of the calcaneus. If necessary, a saw can be used to scrape the bone surfaces smooth to fit the contour of the plate. Attach the lateral Subtalar Precision Guide Arm to the plate prior to provisional fixation, if desired (shown).



Secure the plate to the lateral aspect of the tibiotalar joint using a long Olive Wire in a circular tibial hole and a short Olive Wires in the talus and calcaneus, as shown. Confirm plate position using fluoroscopy.



NOTE: The small hole in the talar portion of the plate is directly in line with the center of the tibial portion of the plate. Talar position can be verified using this hole for guidance. Confirm trajectory of subtalar screw using fluoroscopy.


PERMANENT FIXATION – PLATE SCREWS




NOTE: The talar screw holes accept Ø3.5 mm or Ø4.2 mm non-locking and locking screws. The use of Ø4.2 mm screws is demonstrated in this technique. When using the Ø3.5 mm screws, use the appropriate instrumentation as described on page 4.



Retrieve the Ø4.2 mm Locking Drill Guide and thread into a talar body screw hole. Drill, using the Ø2.8 mm Drill.



Remove the Ø4.2 mm Locking Drill Guide and measure screw length using the Depth Gauge (not shown). Insert the selected screw into the plate hole using the provided Driver and Handle. Do not fully tighten the screw until the second talar screw is secure, to prevent toggling of the plate.




Remove the Olive Wire from the talus. Insert remaining locking or non-locking talar screws, per surgeon preference. It is recommended to drill using the Locking Drill Guide or the straight end of the Easy Guide, in order to achieve on-axis screw trajectories, which allows for placement of a crossing screw via the Precision Guide.



NOTE: Screw placement within the talus may be limited in cases of diseased or eroded tali.

PERMANENT FIXATION – PLATE SCREWS

NOTE: The tibial and calcaneal screw holes accept Ø4.5 mm, Ø4.7 mm, or Ø5.2 mm screws. A laser etched dot on the plate indicates the plate holes that accept the Ø4.5 mm, Ø4.7 mm, and Ø5.2 mm screws. The instructions provided below are for Ø4.5 mm screws. When using the Ø4.7 mm or Ø5.2 mm screws, use the appropriate instrumentation as described on page 4.



Retrieve the Ø4.5 mm oblong Compression Slot Drill Guide and insert into the tibial compression slot with the arrow pointing toward the tibiotalar joint. Drill, using a Ø3.1 mm Drill through the Compression Slot Drill Guide.




Remove the Compression Slot Drill Guide and measure screw length using the Depth Gauge. Insert a Ø4.5 mm non-locking screw but do not fully seat.

PERMANENT FIXATION – TIBIOTALAR PRECISION GUIDED CROSSING SCREW

Retrieve the lateral TT Precision Guide Arm and Set Screw. The Precision Guide is attached to the plate as described on page 9. Place a K-wire Tube into the TT Precision Guide.

NOTE: K-wire Tubes for the Precision Guide are available in Ø1.6 mm and Ø2.3 mm, allowing for Ø5.5 mm or Ø7.0 mm Monster Screws to be used. Partially threaded and fully threaded screw options are available for each screw diameter, per surgeon preference. The use of a Ø7.0 mm Monster Screw is demonstrated in this technique. When using the Ø5.5 mm Monster Screw, use the appropriate instrumentation.




Insert a Ø7.0 mm partially threaded headed Monster Screw across the TT joint using the technique described on pages 10 and 11. Remove the provisional fixation wire across the tibiotalar joint and the tibial and calcaneal Olive Wires prior to fully seating the Monster Screw. Confirm screw length and placement using fluoroscopy. Remove the Ø2.3 mm K-wire serving as a guide wire.




Seat the Ø4.5 mm non-locking screw in the tibia.

PERMANENT FIXATION – SUBTALAR PRECISION GUIDED CROSSING SCREW




Retrieve the K-wire Tube and place into the lateral subtalar Precision Guide Arm. Insert a Ø2.3 mm K-wire through the K-wire Tube across the subtalar joint.



Using the method outlined on pages 14 and 15, place a partially threaded Ø7.0 mm Monster Screw across the subtalar joint. Remove provisional fixation across the subtalar joint prior to fully seating the Monster Screw.


PERMANENT FIXATION – TIBIA AND CALCANEAL PLATE SCREWS



Screw placement in the remaining tibial and calcaneal plate holes is completed at this time using Ø4.5 mm, Ø4.7 mm, or Ø5.2 mm locking or non-locking screws, per surgeon preference. It is recommended to drill using the Locking Drill Guide or the straight end of the Easy Guide, in order to achieve an on-axis screw trajectory.



TIP: If off-axis plate screw trajectories are desired, it is recommended to angle the screws away from the laser markings on the plate, which indicate Precision-guided crossing screw trajectories.



CLOSURE

Proceed to incision closure or concomitant procedures at this time.

INCISION/EXPOSURE

A lateral incision is made over the posterior half of the fibula, beginning approximately 10 cm proximal to the tip of the fibula and curving anterior distally toward the 4th metatarsal, just past the tip of the fibula.

Identify the sural nerve and retract it posteriorly. Continue dissection to the fibula. While retracting the peroneal tendons and sural nerve, a transverse fibular osteotomy is performed by beveling the saw from proximal lateral to distal medial to avoid a sharp bony prominence above the plate. Transect the syndesmotic and lateral ankle ligaments to free the fibula from adjacent soft tissues. Resect the fibula and retain for bone graft, if desired.

Elevate the anterior joint capsule and nearby periosteum to access the anterior tibiotalar joint articulation. Minimal dissection of the talar neck is recommended to avoid devascularization of this bone. Remove any anterior osteophytes that may interfere with joint reduction. Elevate the posterior soft tissues using a periosteal elevator to allow for retractors to be placed anterior to and posterior to the tibiotalar joint.




JOINT PREPARATION

Preparation of the tibiotalar joint can be performed as described on page 16, or per surgeon preference using the provided joint preparation instrumentation.

PERMANENT FIXATION

Steps for fixation of a lateral TT plate, screws and crossing screw can be performed as described previously for the lateral TTC plate.



INCISION/EXPOSURE

TIBIOTALAR JOINT

A prone position should be utilized for this approach, with the foot and ankle extending off the end of the table to allow for a neutral foot position to be achieved. A longitudinal midline, full-thickness incision is made over the posterior tibiotalar joint centered over the Achilles tendon, beginning approximately 10 cm proximal to the ankle joint and terminating at the superior aspect of the calcaneus.

The Achilles tendon can be split along the midline, with care being taken to avoid undermining the subcutaneous tissue and paratenon. The deep fascia is incised and the flexor hallucis longus muscle is retracted medially. Upon exposing the posterior tibia and talus, a capsulotomy is made to expose the posterior tibiotalar joint. Remove any osteophytes on the tibia and talus to allow for exposure to the tibiotalar joint and facilitate entry of instrumentation for cartilage removal.



JOINT PREPARATION

Preparation of the tibiotalar joint can be performed using the provided joint preparation instrumentation. A Hindfoot Distractor is offered to allow for space and visualization during joint preparation, to be used with the Ø2.5 mm K-wires. The Hindfoot Distractor and K-wires are placed through the lateral aspect of the incision with appropriate soft tissue retraction, in a posterior to anterior direction. Following cartilage removal, it is advised to penetrate the subchondral plate with the Subchondral Drill, Burrs and/or Bone Fenestration Chisels to promote healing.

PROVISIONAL FIXATION


Align the ankle joint. The foot and ankle should be positioned such that the ankle is neutral with respect to dorsiflexion and plantarflexion. The foot should be in approximately 5-10° of external rotation and 5° of hindfoot valgus. With the foot and ankle held in this alignment use one or two Ø2.0 mm K-wires to temporarily fixate the tibiotalar joint, per surgeon preference. The provisional fixation wire should be placed in the posterolateral to anteromedial direction for the tibiotalar joint.



PLATE PLACEMENT



Resection of osteophytes or prominent bone may be necessary to ensure proper plate fit. Retrieve the appropriate posterior TT plate based on the patient's anatomy. To position the plate, center the talar portion of the plate on the talus, while ensuring that the proximal plate is midline or just lateral to midline.



Secure the plate to the tibiotalar joint using a long Olive Wire in the most proximal circular hole on the tibia and a long Olive Wire in the medial (shown) or lateral talar screw hole, per surgeon preference. Confirm plate position using fluoroscopy.

PERMANENT FIXATION – PLATE SCREWS



NOTE: The talar screw holes accept Ø3.5 mm or Ø4.2 mm non-locking and locking screws. Ø4.2 mm screws are recommended for this area, except in the case of a small patient. The use of Ø4.2 mm screws is demonstrated in this technique. When using Ø3.5 mm screws, use the appropriate instrumentation as described on page 4.



Retrieve the Ø4.2 mm Locking Drill Guide and thread into the lateral talar body screw hole. Drill, using the Ø2.8 mm Drill.




Remove the Ø4.2 mm Locking Drill Guide and measure screw length using the Depth Gauge. Confirm screw projection and length using the Depth Gauge under fluoroscopy (not shown). Insert the selected screw in the plate using the provided Driver and Handle. Do not fully tighten the talar screws until the tibial compression screw is placed and the Monster Screw is fully seated.



NOTE: The appropriate Locking Drill Guide is recommended for the talar screws and two distal tibia screws to achieve on-axis trajectory.

PERMANENT FIXATION – PLATE SCREWS


Remove the Olive Wire in the medial talar body. Insert a second Ø4.2 mm screw into the medial talar body hole using the same procedure previously described. Do not fully seat.




PERMANENT FIXATION – TIBIAL COMPRESSION SCREW



NOTE: The tibial screw holes accept Ø4.5 mm Ø4.7 mm, or Ø5.2 mm locking or non-locking screws. A laser etched dot on the plate indicates the plate holes that accept Ø4.5 mm, Ø4.7 mm, and Ø5.2 mm screws. The technique demonstrates the use of the Ø4.5 mm screws. When using the Ø4.7 mm or Ø5.2 mm screws, use the appropriate instrumentation as described on page 4.



Retrieve the Ø4.5 mm oblong Compression Slot Drill Guide and insert into the tibial compression slot with the arrow pointing towards the tibiotalar joint. Drill, using a Ø3.1 mm Drill through the Compression Slot Drill Guide.





Remove the Compression Slot Drill Guide and measure screw length using the Depth Gauge. Insert a Ø4.5 mm non-locking screw. Do not fully tighten the screw until the tibiotalar crossing screw has been inserted and the talar body screws have been fully seated. This allows for travel of the compression screw distally in the slot during final tightening of the crossing screw.

PERMANENT FIXATION – PRECISION GUIDED CROSSING SCREW


Assemble and attach the posterior Precision Guide to the posterior TT plate as described on page 9.



NOTE: K-wire Tubes for the Precision Guide are available in Ø1.6 mm and Ø2.3 mm, allowing for Ø5.5 mm or Ø7.0 mm Monster Screws to be used. Partially threaded and fully threaded screw options are available for each screw diameter, per surgeon preference. The use of a Ø7.0 mm Monster Screw is demonstrated in this technique. When using the Ø5.5 mm Monster Screw, use the appropriate instrumentation.




Insert the tibiotalar crossing screw as described on pages 10 and 11. Prior to seating the Monster Screw head against bone, remove the Olive Wire from the tibial screw hole and the provisional fixation across the tibiotalar joint. Fully seat the Monster Screw, then confirm screw length and placement using fluoroscopy. Remove the Ø2.3 mm K-wire.




Fully seat the talar body screws.


PERMANENT FIXATION – PLATE SCREWS



Fully seat the non-locking screw in the tibial compression slot.



Insert the remaining tibial screws using the technique previously described for Ø4.5 mm screws.



Confirm screw position and placement using fluoroscopy.

CLOSURE

Proceed to incision closure or concomitant procedures at this time.

INCISION/EXPOSURE

The incision for the posterior TTC plate is made as described on page 21 for the posterior TT arthrodesis surgical technique; however, a longer incision is made to allow for access to the subtalar joint, in addition to the tibiotalar joint.

A capsulotomy is made to expose the posterior tibiotalar and subtalar joints. Remove any osteophytes on the tibia, talus, and calcaneus to allow for exposure to the tibiotalar and subtalar joints and facilitate entry of instrumentation for cartilage removal. All osteophytes must be removed from the posterior tibiotalar and subtalar joints to facilitate application of the plate.



JOINT PREPARATION

Preparation of the tibiotalar joint can be performed using the provided joint preparation instrumentation. A Hindfoot Distractor is offered to allow for space and visualization during joint preparation, to be used with the Ø2.5 mm K-wires. It is advised to penetrate the subchondral plate with the Subchondral Drill, Burrs and/or Chisels to promote healing.

Remove cartilage from the posterior, middle and anterior facets of the subtalar joint. Perform subchondral plate penetration to these joints to promote healing.

PROVISIONAL FIXATION


Align the ankle joint. The foot and ankle should be positioned such that the ankle is neutral with respect to dorsiflexion and plantarflexion. The foot should be in approximately 5-10° of external rotation and 5° of hindfoot valgus. With the foot and ankle held in this alignment, use multiple Ø2.0 mm K-wires to temporarily fix the tibiotalar and subtalar joints as shown. Provisional fixation wires should be placed in the posterolateral to anteromedial direction for the tibiotalar joint and posterior to anterior direction for the subtalar joint, biased slightly medially.



PLATE POSITIONING AND PROVISIONAL FIXATION



Resection of osteophytes or prominent bone may be necessary to ensure proper plate fit. Retrieve the appropriate posterior TTC plate based on the patient's anatomy. To position the plate, center the talar portion of the plate on the talus, while ensuring that the proximal plate is midline or just lateral to midline and that the distal plate is centered over the superior calcaneus.



Secure the plate to the posterior aspect of the tibiotalar joint using a long Olive Wire in the circular tibial hole inferior to the compression slot and a long Olive Wire in the medial or lateral (shown) talar hole, per surgeon preference. Confirm plate position using fluoroscopy.

PERMANENT FIXATION – PLATE SCREWS




NOTE: The talar screw holes accept Ø3.5 mm or Ø4.2 mm non-locking and locking screws. The use of Ø4.2 mm screws is demonstrated in this technique. When using the Ø3.5 mm screws, use the appropriate instrumentation as describe on page 4.

Retrieve the Ø4.2 mm Locking Drill Guide and thread into the medial talar body screw hole. Drill, using the Ø2.8 mm Drill.





NOTE: The appropriate Locking Drill Guide is recommended for the talar screws, calcaneal screws and two distal tibia screws to achieve on-axis trajectory and avoid interference with the Precision Guided crossing screws.



PERMANENT FIXATION – PLATE SCREWS

Remove the Ø4.2 mm Locking Drill Guide and measure screw length using the Depth Gauge (not shown). Insert the selected screw into the plate hole using the Driver and Handle. Do not fully tighten the talar screws until the Monster Screw is fully seated, to prevent toggling of the plate.



Remove the adjacent Olive Wire from the lateral talar body. Insert the remaining locking or non-locking talar screw, per surgeon preference.




NOTE: The lateral talar body screw must be 24 mm or shorter in length to avoid the projected trajectory of the subtalar crossing screw.



NOTE: The tibial and calcaneal screw holes accept Ø4.5, Ø4.7 mm, or Ø5.2 mm screws. A laser etched dot on the plate indicates the plate holes that accept the Ø4.5 mm, Ø4.7 mm, and Ø5.2 mm screws. The instructions provided below are the Ø4.5 screws. When using the Ø4.7 mm or Ø5.2 mm screws, use the appropriate instrumentation as described on page 4.


Retrieve the Ø4.5 mm oblong Compression Slot Drill Guide and insert into the tibial compression slot with the arrow pointing toward the tibiotalar joint. Drill, using a Ø3.1 mm Drill through the Compression Slot Drill Guide.

Remove the Compression Slot Drill Guide and measure screw length using the Depth Gauge. Insert a Ø4.5 mm non-locking screw, but do not fully seat the compression screw until the tibiotalar crossing screw has been inserted and the talar body screws are fully seated. This allows for travel of the compression screw distally in the slot during final tightening of the crossing screw.




PERMANENT FIXATION – TIBIOTALAR PRECISION GUIDED CROSSING SCREW

Retrieve the posterior TT Precision Guide Arm and Set Screw. The Precision Guide is attached to the plate as described on page 9.




NOTE: K-wire Tubes for the Precision Guide are available in Ø1.6 mm and Ø2.3 mm, allowing for Ø5.5 mm or Ø7.0 mm Monster Screws to be used. Partially threaded and fully threaded screw options are available for each screw diameter, per surgeon preference. The use of a Ø7.0 mm Monster Screw is demonstrated in this technique. When using the Ø5.5 mm Monster Screw, use the appropriate instrumentation.


Insert a Ø7.0 mm partially threaded headed Monster Screw using the technique described on pages 10 and 11. Prior to seating the Monster Screw head against bone, remove the Olive Wire from the tibial screw hole and the provisional fixation across the tibiotalar joint. Fully seat the Monster Screw, then confirm screw length and placement using fluoroscopy. Remove the Ø2.3 mm K-wire serving as a guide wire.



PERMANENT FIXATION – TIBIOTALAR PRECISION GUIDED CROSSING SCREW




Fully seat the talar body screws.




Fully seat the tibial compression screw.

PERMANENT FIXATION – SUBTALAR PRECISION GUIDED CROSSING SCREW



Retrieve the posterior Subtalar Precision Guide Arm and Set Screw. The Precision Guide is attached to the plate as described on page 9. Insert a Ø2.3 mm K-wire through the K-wire Tube across the subtalar joint.




Using the method outlined on pages 10 and 11, insert a partially threaded Ø7.0 mm Monster Screw across the subtalar joint. Remove the provisional fixation wire across the subtalar joint prior to fully seating the Monster Screw.

PERMANENT FIXATION – TIBIA AND CALCANEAL PLATE SCREWS


Screw placement in the remaining tibial and calcaneal plate holes is completed at this time using Ø4.5 mm, Ø4.7 mm, or Ø5.2 mm locking or non-locking screws, per surgeon preference.



TIP: If off-axis plate screw trajectories are desired, it is recommended to angle the screws away from the laser markings on the plate, which indicate Precision-guided crossing screw trajectories.



Confirm screw position and placement using fluoroscopy.



CLOSURE

Proceed to incision closure or concomitant procedures at this time.


Silverback Anterior Mini-Open Plate Caddy

Anterior Mini Open TT plates with corresponding Precision Guide Arms and Precision Guide Arms Base are located within the Mini-Open Anterior Plate Caddy.



Silverback Anterior Plate Caddy

Anterior TT, and TTC plates with corresponding Precision Guide Arms and Precision Guide Arms Base are located within the Anterior Plate Caddy.



Silverback Lateral Plate Caddy

Lateral TT and TTC plates and corresponding Precision Guides are located within the Lateral Plate Caddy.



Silverback K-wire and Olive Wire Caddy

Smooth and threaded K-wires and Olive Wires and a ruler are located within the K-wire and Olive Wire Caddy.




Silverback Posterior Plate Caddy

Posterior TT and TTC plates and corresponding Precision Guides are located within the Posterior Plate Caddy.



Silverback Instrument Tray

All drill guides, drills, overdrills, taps, drivers, forceps and a depth gauge are located within the Silverback



Silverback Screw Caddy

The Silverback screw length options for locking and non-locking screws:


3.5 mm	2 mm increments, 14-30 mm	
4.2 mm	2 mm increments, 14-50 mm	
4.2 mm	5 mm increments, 55-60 mm	
4.5 mm	2 mm increments, 14-50 mm	
4.5 mm	5 mm increments, 55-60 mm	
5.2 mm	2 mm increments, 14-50 mm	
5.2 mm	5 mm increments, 55-60 mm	

The Silverback compact screw length options are as follows:

4.7 mm	2 mm increments, 20-40 mm	
--------	---------------------------	--

Silverback Case Base

Handles, plate bending instrumentation and joint preparation instrumentation including curettes, osteotomes, Chisels and a cartilage removal tool are located at the bottom of the Silverback Instrument Case.



GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - ANTERIOR PLATE CADDY

Part #	Description	Use
P53-401-L101	Anterior TT Plate, Standard, Contoured, Left, 08-Hole, TiA	Single-use
P53-401-R101	Anterior TT Plate, Standard, Contoured, Right, 08-Hole, TiA	Single-use
P53-401-L102	Anterior TT Plate, Long, Contoured, Left, 10-Hole, TiA	Single-use
P53-401-R102	Anterior TT Plate, Long, Contoured, Right, 10-Hole, TiA	Single-use
P53-401-L201	Anterior TT Plate, Standard, Flat, Left, 08-Hole, TiA	Single-use
P53-401-R201	Anterior TT Plate, Standard, Flat, Right, 08-Hole, TiA	Single-use
P53-401-L202	Anterior TT Plate, Long, Flat, Left, 10-Hole, TiA	Single-use
P53-401-R202	Anterior TT Plate, Long, Flat, Right, 10-Hole, TiA	Single-use
P53-401-L111	Anterior TTC Plate, Contoured, Left, 08-Hole, TiA	Single-use
P53-401-R111	Anterior TTC Plate, Contoured, Right, 08-Hole, TiA	Single-use
P53-401-L211	Anterior TTC Plate, Flat, Left, 08-Hole, TiA	Single-use
P53-401-R211	Anterior TTC Plate, Flat, Right, 08-Hole, TiA	Single-use
P51-980-5001	Precision Guide, Arm, Contoured Plates	Reusable
P51-980-5101	Precision Guide, Arm, Flat Plates	Reusable
P51-980-5003	Precision Guide Set Screw, M3, Threaded	Reusable
P51-980-5005	Anterior Plates, Ball Tip Wire Guide	Reusable

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - MINI-OPEN PLATE CADDY

Part #	Description	Use
P53-431-L101	Anterior Mini-Open Plate, TT, Contoured, Standard, Left	Single-use
P53-431-R101	Anterior Mini-Open Plate, TT, Contoured, Standard	Single-use
P53-431-L201	Anterior Mini-Open Plate, TT, Flat, Standard, Left	Single-use
P53-431-R201	Anterior Mini-Open Plate, TT, Flat, Standard, Right	Single-use
P51-983-5001	Precision® Guide Arm, Mini-Open, Tibiotalar Joint, Medial	Reusable
P51-983-5002	Precision® Guide Arm, Mini-Open, Tibiotalar Joint, Lateral	Reusable
P51-983-5003	Precision® Guide Threaded Tower, Mini-Open	Reusable
P51-983-5004	Precision® Guide Arm Base, Mini-Open, Contoured	Reusable
P51-983-5005	Precision® Guide Arm Base, Mini-Open, Flat	Reusable
P51-980-5003	Precision Guide Set Screw, SILVERBACK	Reusable
P51-980-5010-16	Precision Guide K-Wire Sleeve, 1.6m	Reusable
P51-980-5010-23	Precision Guide K-Wire Sleeve, 2.3m	Reusable
P99-250-1610	Olive Wire, SMOOTH, 1.6 x 100mm, Tower Wire, X-Long	Single-use
P99-251-1610	Olive Wire, THREADED, 1.6 x 100mm, Tower Wire, X-Long	Single-use
P99-250-1611	Olive Wire, SMOOTH, 1.6 x 110mm, Tower Wire, X-Long	Single-use
P99-251-1611	Olive Wire, THREADED, 1.6 x 110mm, Tower Wire, X-Long	Single-use

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - LATERAL PLATE CADDY

Part #	Description	Use
P53-402-L001	Lateral TT Plate, Standard, Left, 13-Hole, TiA	Single-use
P53-402-R001	Lateral TT Plate, Standard, Right, 13-Hole, TiA	Single-use
P53-402-L101	Lateral TTC Plate, Standard, Left, 17-Hole, TiA	Single-use
P53-402-R101	Lateral TTC Plate, Standard, Right, 17-Hole, TiA	Single-use
P53-402-L102	Lateral TTC Plate, Long, Left, 19-Hole, TiA	Single-use
P53-402-R102	Lateral TTC Plate, Long, Right, 19-Hole, TiA	Single-use
P51-981-5001	Precision Guide, Tibiotalar Joint, Guide Arm	Reusable
P51-981-510R	Precision Guide, Subtalar Joint, Guide Arm, Right	Reusable
P51-981-510L	Precision Guide, Subtalar Joint, Guide Arm, Left	Reusable
P51-980-5003	Precision Guide Set Screw, M3, Threaded	Reusable

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - POSTERIOR PLATE CADDY

Part #	Description	Use
P53-403-L101	Posterior TT Plate, Standard, Contoured, Left, 06 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L102	Posterior TT Plate, Long, Contoured, Left, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R101	Posterior TT Plate, Standard, Contoured, Right, 06 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R102	Posterior TT Plate, Long, Contoured, Right, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L201	Posterior TT Plate, Standard, Flat, Left, 06 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L202	Posterior TT Plate, Long, Flat, Left, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R201	Posterior TT Plate, Standard, Flat, Right, 06 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R202	Posterior TT Plate, Long, Flat, Right, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L111	Posterior TTC Plate, Standard, Contoured, Left, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L112	Posterior TTC Plate, Long, Contoured, Left, 10 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R111	Posterior TTC Plate, Standard, Contoured, Right, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R112	Posterior TTC Plate, Long, Contoured, Right, 10 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L211	Posterior TTC Plate, Standard, Flat, Left, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-L212	Posterior TTC Plate, Long, Flat, Left, 10 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R211	Posterior TTC Plate, Standard, Flat, Right, 08 $\frac{1}{8}$ " Hole, TiA	Single-use
P53-403-R212	Posterior TTC Plate, Long, Flat, Right, 10 $\frac{1}{8}$ " Hole, TiA	Single-use
P51-982-5001	Precision Guide, Tibiotalar Joint, Guide Arm	Reusable
P51-982-5101	Precision Guide, Subtalar Joint, Guide Arm	Reusable
P51-980-5003	Precision Guide Set Screw, M3, Threaded	Reusable

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - WIRE CADDY

Part #	Description	Use
P99-150-0005	K-Wire Hole Gauge & Ruler	Reusable
P99-192-2020	K-Wire, Single Ended Trocar Tip, SMOOTH, 2.0 x 200mm	Single-use
P99-193-2020	K-Wire, Single Ended Trocar Tip, Threaded, 2.0 x 200mm	Single-use
P99-192-2515	K-Wire, Single Ended Trocar Tip, SMOOTH, 2.5 x 150mm	Single-use
P99-193-2515	K-Wire, Single Ended Trocar Tip, Threaded, 2.5 x 150mm	Single-use
P99-204-1608	Half Olive Wire, Large Olive (7mm), ø1.60mm x 8cm, Smooth	Single-use
P99-205-1608	Half Olive Wire, Large Olive (7mm), ø1.60mm x 8cm, Threaded	Single-use
P99-204-1609	Half Olive Wire, Large Olive (7mm), ø1.60mm x 9cm, Smooth	Single-use
P99-205-1609	Half Olive Wire, Large Olive (7mm), ø1.60mm x 9cm, Threaded	Single-use

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - CASE INSTRUMENTS

Part #	Description	Use
P99-900-1005	P28, Instrument, FootPlate	Reusable
P99-000-AOLG	Mini-AO, Ratcheting Handle Streamline, Cannulated, Large	Reusable
P99-000-AOMN	Mini-AO, Ratcheting Handle Streamline, Cannulated	Reusable
P99-000-316T	3/16" SQ, Ratcheting T-Handle, Cannulated	Reusable
P99-150-0010	Hindfoot Distractor, 2.5 & 3.0 K-Wire	Reusable
P51-910-1002	Flat Plate Bender	Reusable
P51-910-5001	Threaded Bending Bar, Double Ended, Silverback/R3CON	Reusable
P99-150-0055	Cartilage Removal Tool, Large	Reusable
P99-150-0094	Curette, 20° Angled, 7mm	Reusable
P99-150-0095	Curette, Ring, Straight, 7mm	Reusable
P99-150-0096	Curette, Curved, 7mm	Reusable
P99-150-0097	Curette, Ring, 20° Angled, 7mm	Reusable
P99-150-1241	Osteotome, Straight, Handled, 6mm	Reusable
P99-150-1243	Osteotome, Straight, Handled, 12mm	Reusable
P99-150-1340	Osteotome, Curved, Handled, 3mm	Reusable
P99-150-1341	Osteotome, Curved, Handled, 6mm	Reusable
P99-150-1343	Osteotome, Curved, Handled, 12mm	Reusable
P99-150-0035	Bone Fenestration Chisel	Reusable
P99-150-0135	Bone Fenestration Chisel, Curved	Reusable

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - INSTRUMENT CADDY

Part #	Description	Use
P99-150-0136	Depth Gauge Plate Screw, 85mm Screw	Reusable
P51-900-1005	3.5 Locking Drill Guide, Long, R3CON™	Reusable
P51-900-1006	4.2 Locking Drill Guide, Long, R3CON™	Reusable
P51-900-5005	4.5 Locking Drill Guide, Long, Silverback	Reusable
P51-900-5006	5.2 Locking Drill Guide, Long, Silverback	Reusable
P51-901-5135	Easy Guide, Long, Double-sided, Cone/Straight, 3.5mm	Reusable
P51-901-5142	Easy Guide, Long, Double-sided, Cone/Straight, 4.2mm	Reusable
P51-901-5145	Easy Guide, Long, Double-sided, Cone/Straight, 4.5mm	Reusable
P51-901-5152	Easy Guide, Long, Double-sided, Cone/Straight, 5.2mm	Reusable
P51-902-1005	Centering Drill Guide, Long, 3.5mm, R3CON™	Reusable
P51-902-1006	Centering Drill Guide, Long, 4.2mm, R3CON™	Reusable
P51-902-5005	Centering Drill Guide, Long, 4.5mm, Silverback	Reusable
P51-902-5006	Centering Drill Guide, Long, 5.2mm, Silverback	Reusable
P51-905-5001	Over Drill Guide, Double Ended (3.5/4.2 & 4.5/5.2mm)	Reusable
P51-903-5005	4.5mm Compression Slot Drill Guide, Silverback	Reusable
P51-903-5006	5.2mm Compression Slot Drill Guide, Silverback	Reusable
P99-100-2416	Drill, 2.4 x 160mm, Solid, Measuring, Long, AO	Reusable
P99-100-2816	Drill, 2.8 x 160mm, Solid, Measuring, Long, AO	Reusable
P99-100-3116	Drill, 3.1 x 160mm, Solid, Measuring, Long, AO	Reusable
P99-100-3616	Drill, 3.6 x 160mm, Solid, Measuring, Long, AO	Reusable
P99-100-3511	Over Drill, 3.5 x 110mm, Solid, AO	Reusable
P99-100-4212	Over Drill, 4.2 x 120mm, Solid, AO	Reusable
P99-100-4512	Over Drill, 4.5 x 120mm, Solid, AO	Reusable
P99-100-5212	Over Drill, 5.2 x 120mm, Solid, AO	Reusable
P99-100-2010	Bone Fenestration Perforator	Reusable
P99-150-0001	P28, Screw Forceps, Ti	Reusable
P51-911-4500	Tap, Double Lead, ø4.5mm	Reusable
P51-911-4700	Tap, Single Lead, ø4.7mm	Reusable
P51-911-5200	Tap, Double Lead, ø5.2mm	Reusable
P99-191-LR10	Screw Driver Attachment, Solid, Long, AO, HX-10, Short Taper, R3CON™	Reusable
P99-191-SB15	Screw Driver Attachment, Solid, Long, AO, HX-15, Short Taper, Silverback	Reusable

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - SLIM LINE IMPLANT CADDY

Part #	Description	Use
P50-353-35[14-30]	R3Con Locking Plate Screw, 3.5 X 14-30 mm (2 mm increments)	Single-Use
P50-453-35[14-30]	R3Con Non-Locking Plate Screw, 3.5 X 14-30 mm (2 mm increments)	Single-Use
P50-353-42[14-40]	R3Con Locking Plate Screw, 4.2 X 14-40 mm (2 mm increments)	Single-Use
P50-453-42[14-40]	R3Con Non-Locking Plate Screw, 4.2 X 14-40 mm (2 mm increments)	Single-Use
P50-553-45[14-60]	Silverback Locking Plate Screw, 4.5 X 14-48 mm (2 mm increments) 50-60 mm (5 mm increments)	Single-Use
P50-553-52[14-60]	Silverback Locking Plate Screw, 5.2 X 14-48 mm (2 mm increments) 50-60 mm (5 mm increments)	Single-Use
P50-653-45[14-60]	Silverback Non-Locking Plate Screw, 4.5 X 14-48 mm (2 mm increments) 50-60 mm (5 mm increments)	Single-Use
P50-653-52[14-60]	Silverback Non-Locking Plate Screw, 5.2 X 14-48 mm (2 mm increments) 50-60 mm (5 mm increments)	Single-Use
P53-401-L101	Anterior Tt Plate, Standard, Contoured, Left	Single-Use
P53-401-L102	Anterior Tt Plate, Long, Contoured, Left	Single-Use
P53-401-L111	Anterior Ttc Plate, Contoured, Left	Single-Use
P53-401-L201	Anterior Tt Plate, Standard, Flat, Left	Single-Use
P53-401-L202	Anterior Tt Plate, Long, Flat, Left	Single-Use
P53-401-L211	Anterior Ttc Plate, Flat, Left	Single-Use
P53-401-R101	Anterior Tt Plate, Standard, Right	Single-Use
P53-401-R102	Anterior Tt Plate, Long, Right	Single-Use
P53-401-R111	Anterior Ttc Plate, Contoured, Right	Single-Use
P53-401-R201	Anterior Tt Plate, Standard, Flat, Right	Single-Use
P53-401-R202	Anterior Tt Plate, Long, Flat, Right	Single-Use
P53-401-R211	Anterior Ttc Plate, Flat, Right	Single-Use
P53-431-L101	Anterior Mini-Open Plate, Tt, Contoured, Standard, Left, 6 Hole, TIA	Single-Use
P53-431-L201	Anterior Mini-Open Plate, Tt, Flat, Standard, Left, 6 Hole, TIA	Single-Use
P53-431-R101	Anterior Mini-Open Plate, Tt, Contoured, Standard, Right, 6 Hole, TIA	Single-Use
P53-431-R201	Anterior Mini-Open Plate, Tt, Flat, Standard, Right, 6 Hole, TIA	Single-Use

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - SLIM LINE INSTRUMENT CADDY

Part #	Description	Use
P51-900-1005	3.5 Locking Drill Guide, Long, R3Con™	Reusable
P51-900-1006	4.2 Locking Drill Guide, Long, R3Con™	Reusable
P51-900-5005	4.5 Locking Drill Guide, Long, Silverback	Reusable
P51-900-5006	5.2 Locking Drill Guide, Long, Silverback	Reusable
P51-901-5135	Easy Guide, Long, Double-Sided, Cone/Straight, 3.5mm	Reusable
P51-901-5142	Easy Guide, Long, Double-Sided, Cone/Straight, 4.2mm	Reusable
P51-901-5145	Easy Guide, Long, Double-Sided, Cone/Straight, 4.5mm	Reusable
P51-901-5152	Easy Guide, Long, Double-Sided, Cone/Straight, 5.2mm	Reusable

GORILLA SILVERBACK ANKLE FUSION PLATING SYSTEM - SLIM LINE INSTRUMENT CADDY CONTINUED

Part #	Description	Use
P51-903-5005	4.5mm Compression Slot Drill Guide, Silverback	Reusable
P51-903-5006	5.2mm Compression Slot Drill Guide, Silverback	Reusable
P51-910-1002	Flat Plate Bender	Reusable
P51-910-5001	Threaded Bending Bar, Double Ended, Silverback	Reusable
P99-000-AOLG	Mini-Ao, Ratcheting Handle Streamline, Cannulated, Large	Reusable
P99-100-2010	Bone Fenestration Perforator	Reusable
P99-100-2416	Drill, 2.4 X 160mm, Solid, Measuring, Long, Ao	Reusable
P99-100-2816	Drill, 2.8 X 160mm, Solid, Measuring, Long, Ao	Reusable
P99-100-3116	Drill, 3.1 X 160mm, Solid, Measuring, Long, Ao	Reusable
P99-100-3616	Drill, 3.6 X 160mm, Solid, Measuring, Long, Ao	Reusable
P99-150-0001	Screw Forceps	Reusable
P99-150-0035	Bone Fenestration Chisel	Reusable
P99-150-0055	Cartilage Removal Tool, Large	Reusable
P99-150-0096	Curette, Curved, 7mm	Reusable
P99-150-0097	Curette, Ring, 20° Angled, 7mm	Reusable
P99-150-0135	Bone Fenestration Chisel, Curved	Reusable
P99-150-0136	Depth Gauge Plate Screw, 85mm Screw	Reusable
P99-150-1343	Osteotome, Curved W/Handle, 12mm	Reusable
P99-191-LR10	Screw Driver Attachment, Solid, Long, Ao, Hx-10, Short Taper, R3Con™	Reusable
P99-191-SB15	Screw Driver Attachment, Solid, Long, Ao, Hx-15, Short Taper, Silverback™	Reusable
P51-980-5001	Precision® Guide Arm	Reusable
P51-980-5005	Anterior Plates, Ball Tip Wire Guide	Reusable
P51-980-5101	Precision® Guide Arm, Flat Plates	Reusable
P51-980-5003	Precision® Guide Set Screw, Silverback	Reusable
P51-980-5010-16	Precision® Guide K-Wire Tube, 1.6mm	Reusable
P51-980-5010-23	Precision® Guide K-Wire Tube, 2.3mm	Reusable
P51-983-5001	Anterior Mini-Open Plate, Medial Precision Guide, Tibiotalar Joint	Reusable
P51-983-5002	Anterior Mini-Open Plate, Lateral Precision Guide, Tibiotalar Joint	Reusable
P51-983-5003	Anterior Mini-Open Plate, Joystick	Reusable
P51-983-5004	Anterior Mini-Open Plate, Precision Guide Arm Base, Contoured	Reusable
P51-983-5005	Anterior Mini-Open Plate, Precision Guide Arm Base, Flat	Reusable
P99-250-1610	Olive Wire, Smooth, 1.6 X 100mm, Tower Wire, Long	Single-use
P99-250-1611	Olive Wire, Smooth, 1.6 X 110mm, Tower Wire, X-Long	Single-use
P99-251-1610	1.60mm X 10Cm Olive Wire, 316 Lvm	Single-use
P99-251-1611	Olive Wire, Threaded, 1.6 X 110mm, Tower Wire, X-Long	Single-use
P99-150-0005	K-Wire Hole Gauge & Ruler	Reusable
P99-192-2020	K-Wire, Single Ended Trocar Tip, Smooth, 2.0 X 200mm	Single-use
P99-204-1608	Half Olive Wire, Large Olive (7mm), 1.6 X 80mm, Smooth	Single-use
P99-205-1608	Half Olive Wire, Large Olive (7mm), 1.6 X 80mm, Threaded	Single-use

Refer to www.paragon28.com/ifus for the complete and most current instructions for use document.

INSTRUCTIONS FOR USE: BABY GORILLA®/GORILLA® PLATING SYSTEM

Indications, Contraindications, Warnings and Precautions relevant to the Baby Gorilla®/Gorilla® Plating System are contained in the Instructions for Use document of the Baby Gorilla®/Gorilla® Plating System P51-IFU-3001.

A patient with the Paragon 28® Baby Gorilla®/Gorilla® Plating System may be safely scanned under the following conditions. Failure to follow these conditions may result in injury to the patient.	
Name/Identification of Device	Paragon 28® Baby Gorilla®/Gorilla® Plating System
Nominal value(s) of Static Magnetic Field [T]	1.5 T or 3 T
Maximum Spatial Field Gradient [T/m and gauss/cm]	30 T/m (3000 gauss/cm)
RF Excitation	Circularly Polarized (CP)
RF Transmit Coil Type	Whole body transmit coil, Head RF transmit-receive coil
Operating Mode	Normal Operating Mode
Maximum Whole Body SAR [W/kg]	2.0 W/kg (Normal Operating Mode)
Limits on Scan Duration	2.0 W/kg whole body average SAR for 60 minutes of continuous RF (a sequence or back to back series/scan without breaks)
MR Image Artifact	The presence of this implant may produce an image artifact
If information about a specific parameter is not included, there are no conditions associated with that parameter.	

Refer to www.paragon28.com/ifus for the complete and most current instructions for use document.

INSTRUCTIONS FOR USE: MONSTER®/MINI-MONSTER® SCREW SYSTEM

Indications, Contraindications, Warnings and Precautions relevant to the Monster® Screw System are contained in the Instructions for Use document of the Monster® Screw System P20-IFU-3001.



A patient with the Paragon 28® Monster® Screw System may be safely scanned under the following conditions. Failure to follow these conditions may result in injury to the patient.

Name/Identification of Device	Paragon 28® Monster® Screw System
Nominal value(s) of Static Magnetic Field [T]	1.5 T or 3 T
Maximum Spatial Field Gradient [T/m and gauss/cm]	30 T/m (3000 gauss/cm)
RF Excitation	Circularly Polarized (CP)
RF Transmit Coil Type	Whole body transmit coil, Head RF transmit-receive coil
Maximum Whole Body SAR [W/kg]	2.0 W/kg (Normal Operating Mode)
Limits on Scan Duration	All anatomical regions can be safely scanned under the following conditions: 2.0 W/kg whole body average SAR for 5 minutes of continuous RF (a sequence or back to back series/scan without breaks) with a 20 minute cooling period between scans for an hour long scanning session
	Scanning of the knees and all anatomy superior to the knees can be safely scanned under the following conditions: 2.0 W/kg whole body average SAR for 60 minutes of continuous RF (a sequence or back to back series/scan without breaks)
MR Image Artifact	The presence of this implant may produce an image artifact of 20 mm
If information about a specific parameter is not included, there are no conditions associated with that parameter.	

NOTES

[illegible]



PATENTED, DESIGNED & EXCLUSIVELY DISTRIBUTED BY



P51-STG-3011 RevC [2026-04-20]

© Copyright 2026 Zimmer Biomet®. All rights reserved.

Patents: www.paragon28.com/patents

Paragon 28, Inc. 
14445 Grasslands Dr.
Englewood, CO 80112 USA
(855) 786-2828

Australian Sponsor
Zimmer Biomet Pty Ltd
(A.C.N: 096 480 992)(A.B.N: 96 096 480 992)
Level 4, 12 Narabang Way
BELROSE NSW 2085
AUSTRALIA
Phone: (02) 9483 5400

All content herein is protected by copyright, trademarks and other intellectual property rights, as applicable, owned by or licensed to Zimmer Biomet or its affiliates unless otherwise indicated, and must not be redistributed, duplicated or disclosed, in whole or in part, without the express written consent of Zimmer Biomet.

DISCLAIMER

Zimmer Biomet does not practice medicine. This technique was developed in conjunction with healthcare professionals. This document is intended for surgeons and is not intended for laypersons. Each surgeon should exercise his or her own independent judgment in the diagnosis and treatment of an individual patient, and this information does not purport to replace the comprehensive training surgeons have received. As with all surgical procedures, the product/technique used in each case will depend on the surgeon's medical judgment as the best treatment for each patient. Results will vary based on health, weight, activity and other variables. Not all patients are candidates for this product and/or procedure. Caution: Federal (USA) law restricts this device to sale by or on the order of a surgeon. Rx only. ©2025 Zimmer Biomet.

All content herein is protected by copyright, trademarks and other intellectual property rights, as applicable, owned by or licensed to Zimmer Biomet or its affiliates unless otherwise indicated, and must not be redistributed, duplicated or disclosed, in whole or in part, without the express written consent of Zimmer Biomet.

For indications, contraindications, warnings, precautions, potential adverse effects and patient counseling information, see the instructions for use or contact your local representative; visit www.paragon28.com for additional product information.

This material is intended for health care professionals. Distribution to any other recipient is prohibited.

www.Paragon28.com