


Gorilla[®] Lapidus Arthrodesis Plating System







LAPIDUS PLATE FEATURES

- ► Anatomically contoured to 1st metatarsal and medial cuneiform
- Patented plate shape with plantar arm helps to reduce plantar gapping (excludes Armless Standard plate)
- Medial wall fixation allows for improved resistance to bending compared to dorsal and dorsal medial plates due to the placement of the plate on its side¹
- Ramped proximal portion of plate designed to help reduce irritation to the tibialis anterior
- Medial wall plate curvature aids in limiting adductory forces on 1st metatarsal by supporting corrected position during distal screw insertion



Biconvex curvature aids in minimizing adductory forces while ensuring plate fit to the anatomy.

PRECISION[™] GUIDE LAPIDUS

- Patent pending guide allows for placement of cross-screw across the arthrodesis while maintaining on-axis placement of plate screws
- Provides four trajectories of guide wire paths for variations among patient anatomy
- ▶ Precision[™] Guide Lapidus allows for Ø3.5 mm, Ø4.0 mm or Ø4.5 mm cannulated cross-screw



LAPIDUS PLATES (20 TOTAL PLATES)

	STANDARD ARMLESS	STANDARD	MEDIAL WALL STEP-OFF	GRAFT SPANNING
	000			
Size	Standard	Standard	Step off sizes in 1 mm increments from 1–5 mm	Small (5 mm Graft) Medium (8 mm/10 mm Graft) Large (12 mm/14 mm Graft)
Plate Thickness	1.3 mm	1.3 mm	1.4 mm	1.6 mm
Total Plates	2	2	10	6



PRESERVE[™] LAPIDIUS GRAFT

- Patented shape features both dorsal to plantar and medial to lateral taper allowing for biplanar correction
- Donor harvest site is density matched specific to Lapidus indication for strength demands and blood flow requirements
- Aseptically processed without gamma irradiation or hydrogen peroxide preserve the native mechanical advantages of human bone and the osteoinductivity of the environment in which the graft is being implanted







Allows surgeon to determine ideal graft size by demonstrating correction

LAPIDIUS NIPPER

- Patent pending instrument specifically designed to aid in removal of the two bone fragments created after sagital saw cartilage resection of the 1st TMT Joint
- Osteotome jaw designed to aid in completion of saw cut
- Sharp toothed jaw helps to release remaining soft tissue attachments
- Osteotome jaw and sharp toothed jaw clamp together allowing for extraction of fragment with less disruption of surrounding tissue
- Long jaws designed to grasp around entire bone fragment from dorsal to plantar







Gorilla® Lapidus Arthrodesis Plating System

LAPD-01 Rev E 2023-09-29

Paragon 28°, Inc. **d** 14445 Grasslands Dr. Englewood, CO 80112 USA (855) 786-2828 Paragon 28[®] Medical Devices Trading Limited First Floor Block 7 Beckett Way Park West Business Park Dublin 12 D12 X884 Ireland +353 (0) 1588 0350

[™]Trademarks and [®]Registered Marks of Paragon 28[®], Inc. © Copyright 2023 Paragon 28[®], Inc. All rights reserved. Patents: www.paragon28.com/patents



www.Paragon28.com

For the contraindications, potential complications and adverse reactions, warnings and precautions associated with this device, please refer to the device specific instructions for use at http://www.paragon28.com/ifus