

ANTERIOR & POSTERIOR
OFFSET POLY TRIALS



PRODUCT INFORMATION

The Paragon 28® APEX 3D™ Total Ankle Replacement System is a cemented, fixed bearing device comprised of a tibial component, a talar component and a Vitamin E Ultra-High Molecular Weight Polyethylene component. Implants are available in varying sizes and design configurations intended for both primary and revision applications. For additional information regarding Indications for Use, Contraindications, Warnings, Precautions, etc. please visit: <https://www.paragon28.com/ifus/>

CONTENTS

SECTION 1

Situational Overview

Anterior & Posterior Subluxed Talus.....2

SECTION 2

Auxiliary Surgical Technique: Offset Poly Trial Evaluation

Instrument & Trial Overview.....3

Anterior – Or – Posterior Offset Poly Trial Evaluation.....4

APPENDIX A

Instruction for Use (IFU) & Contraindications, Warnings, Precautions.....5-6

DESIGN TEAM

Mark Dalton, MD – Austin, TX

Jeffrey Christensen, DPM – Everett, WA

Michael Houghton, MD – Fort Collins, CO

Mark Myerson, MD – Denver, CO

ACKNOWLEDGMENTS:

Contributing Surgeon Advisors, Paragon 28's Development Engineers, Clinical Researchers and Marketing Teams.

ANTERIOR & POSTERIOR SUBLUXED TALUS

To accommodate patients with either an anterior or posterior subluxated talus, the APEX 3D™ Total Ankle Replacement System introduced uniquely designed biased poly insert options to shift the tibial insert articulation to accommodate the center of rotation of the subluxed talus.


Malalignment in the sagittal plane may potentially lead to edge loading and contact stresses of the poly. The biased tibial insert allows the talus to maintain its reduced position.



Posteriorly
Subluxed Position





Anteriorly
Subluxed Position



Offset Poly Insert Size Options	1	2	3	4	5
Anterior or Posterior Offset (mm)	3.00	3.25	3.50	3.75	4.00
Offset is distance from the center of the tibial base					

INSTRUMENT OVERVIEW: ANTERIOR & POSTERIOR OFFSET POLY TRIALS




OFFSET POLY TRIAL EVALUATION: ANTERIOR – OR – POSTERIOR SUBLUXED TALUS

- If it is determined through a lateral fluoroscopic evaluation of the Neutral Poly Trial that the patient may benefit from the use of an Anterior or Posterior Offset Poly Insert, the Offset Poly Trial can be used to evaluate and make a final determination.
- Based on the anatomic position of the talus, attach the Offset Poly Trial in either the anterior or posterior orientation to the Poly Trial Handling Tool and insert into the joint such that the dovetail of the Poly Trial connects with the dovetail of the Tibial Trial.
- Put the tibiotalar joint through gentle range of motion evaluation to ensure adequate placement and correct Poly Trial thickness.
- Confirm that the Talar and Poly Trial components are size matched.
- Utilizing a lateral fluoroscopic view, confirm both the Tibial and Talar Trial alignment Markers are in line with their associated Offset Poly Trial alignment markers.



SURGICAL NOTE: Ensure the fluoroscopic notch on the talar trial is visible on lateral fluoroscopy to ensure an optimal fit. Poly trial is equipped with M/L fluoroscopy markers to help determine fit and placement.



Refer to www.paragon28.com/ifus for the complete and most current instructions for use document.

INDICATIONS FOR USE

The APEX 3D™ Total Ankle Replacement System is indicated as a total ankle replacement in primary surgery for patients with ankle joints damaged by severe rheumatoid, post-traumatic, or degenerative arthritis. Revision surgery for these patients is also indicated for patients with sufficient bone stock present. In the United States, components are intended for cemented use only.

CONTRAINDICATIONS

Use of the APEX 3D™ Total Ankle Replacement System is contraindicated in cases of inflammation, cases of active or suspected sepsis/infection and osteomyelitis; or in patients with certain metabolic diseases.

All applications that are not defined by the indications are contraindicated. In addition, surgical success can be adversely affected by:

- Acute or chronic infections, local or systemic
- Any mental or neuromuscular disorder that could result in an unacceptable risk of failure at the time of fixation or complications in post-operative treatment (e.g. dementia, senility, alcoholism)
- Corpulence; an overweight or corpulent patient can strain the prosthesis to such a degree that stabilization or prosthesis failure can occur
- Excessive loads as caused by activity or patient weight
- Female of childbearing age, for whom a negative pregnancy test is not obtained
- Steroid use
- Inadequate neuromuscular status (e.g. prior paralysis, neuropathy, neuropathic joint, fusion and/or inadequate abductor strength)
- Muscular atrophy
- Osteomyelitis
- Poor bone stock, poor skin coverage, or excessive bone loss around the joint which would make the procedure unjustifiable
- Sepsis
- Skeletally immature patients (patient is less than 21 years of age at the time of surgery)
- Suspected or documented metal allergy or intolerance
- Musculoskeletal disease that may adversely affect gait or weightbearing
- Neurologic disorder/instability and non-compliance that may adversely affect gait or weight bearing
- Vascular deficiency in the ankle joint

Other medical or surgical pre-conditions that could compromise the potentially beneficial procedure, such as:

- Congenital abnormalities

- Immunosuppressive pathologies
- Increased sedimentation rates that cannot be explained by other pathologies
- Marked bone loss, severe osteoporosis, or revision procedures for which an adequate fit of the prosthesis cannot be achieved
- Metabolic disorders that may impair bone formation
- Osteomalacia
- Poor prognosis for good wound healing
- Presence of tumors
- Increased leukocyte (WBC) count
- Pronounced left shift in the differential leukocyte count
- Uncooperative patient or patient with neurological disorders, incapable of following instructions

POTENTIAL COMPLICATIONS AND ADVERSE REACTIONS

IN ANY SURGICAL PROCEDURE, THE POTENTIAL FOR COMPLICATIONS AND ADVERSE REACTIONS EXIST. THE RISKS AND COMPLICATIONS WITH THESE PROSTHETIC COMPONENTS INCLUDE:

- Asymptomatic, progressive bone resorption (osteolysis) due to foreign body reaction to particulate matter (See Important Physician Information section for more information)
- Sensitivity, allergy or other reactions to prosthetic component materials
- Peripheral neuropathies or nerve damage resulting in pain or numbness of the affected limb
- Loosening or migration of the prosthetic components
- Subluxation or dislocation of the prosthetic components with resulting reduction in range of movement
- Bending, disassembly and/or breakage of the prosthetic components
- Fractures resulting from unilateral joint loading
- Fatigue fracture of the prosthetic components as the result of trauma, strenuous activity, improper alignment, incomplete implant seating, or duration of service
- Bone fracture by trauma or excessive loading, particularly in the presence of poor bone stock
- Drop in blood pressure intra-operatively due to the use of bone cement
- Thrombosis, embolism, or myocardial infarction
- Wound hematoma and delayed wound healing
- Acute post-operative wound infections and late infections with possible sepsis
- Pain, a feeling of malaise or abnormal sensations due to the prosthetic components

POTENTIAL COMPLICATIONS AND ADVERSE REACTIONS (CONTINUED)

- Inadequate range of motion due to improper selection or positioning of components or periarticular calcification
- Temporary and protracted functional neurological perturbation
- Corrosion with localized tissue reaction and pain
- Bone loss due to stress shielding
- Secondary necrosis of the talus

All possible complications listed here are not typical of Paragon 28®, Inc. products but in principle, may be observed with any total joint replacement implant. Promptly inform Paragon 28®, Inc. as soon as complications occur in connection with the implants or surgical instruments used. In the event of premature failure of an implant in which a causal relationship with its geometry, surface quality or mechanical stability is suspected, please provide Paragon 28®, Inc. with the explant(s) in a cleaned, disinfected and sterile condition. Paragon 28®, Inc. cannot accept any other returns of used implants. The surgeon is held liable for complications associated with inadequate asepsis, inadequate preparation of the osseous implant bed in the case of implants, incorrect indication or surgical technique or incorrect patient information and consequent incorrect patient behavior

WARNINGS AND PRECAUTIONS

- This device is not intended for subtalar joint fusion or subtalar joint impingement. Please carefully evaluate the anatomy of each patient before implantation.
- The surgeon should discuss with the patient prior to surgery possible risks, precautions, warnings, consequences, complications, and adverse reactions associated with the surgical procedure and implantation of the device.
- Improper selection, placement, positioning, and fixation of the prosthetic components may result in unusual stress conditions and a subsequent reduction in service life of the prosthetic component.
- Periodic, long-term follow-up is recommended to monitor the position and state of the prosthetic components, as well as the condition of the adjoining bone.
- Re-operation to remove or replace prosthetic components may be required at any time due to medical reasons or device failure. If corrective action is not taken, complications may occur.
- Patients need to be informed regarding expectations pertaining to performance and limitations following surgery. The prosthesis does not replace normal bone, has a finite service life, and future revision surgeries may be necessary. Protection of the prosthesis from full weight bearing is needed until adequate fixation and healing is achieved. Certain activities and loading trauma should be limited to prevent unreasonable stresses that could lead to breaking or damage of the prosthetic components.

- Do not attempt a surgical procedure with faulty, damaged or suspect instruments or implants. Inspect all components preoperatively to assure utility.
- Never modify an implant.
- The implants and guide wires are intended for single use only.
- Instruments and implants are to be treated as sharps.
- Do not implant the instruments.
- **Do not use other manufacturer's instruments or implants in conjunction with the APEX 3D™ Total Ankle Replacement Device.**
- **Do not re-sterilize the APEX 3D™ Total Ankle Replacement Implants or Instruments.**

IMPORTANT PHYSICIAN INFORMATION

BONE RESORPTION IS A NATURAL CONSEQUENCE OF TOTAL JOINT ARTHROPLASTY DUE TO CHANGES IN BONE REMODELING PATTERNS. BONE REMODELING IS MEDIATED BY THE CHANGES IN STRESS DISTRIBUTION CAUSED BY IMPLANTATION. EXTENSIVE RESORPTION AROUND THE PROSTHESIS MAY LEAD TO IMPLANT LOOSENING AND FAILURE. IT IS GENERALLY AGREED THAT OSTEOLYSIS IS THE RESULT OF LOCALIZED FOREIGN-BODY REACTION TO PARTICULATE DEBRIS GENERATED BY CEMENT, METAL, UHMWPE, AND CERAMIC. REGARDING THE ETIOLOGY, IT HAS BEEN HYPOTHESIZED THAT PARTICULATE DEBRIS GENERATED BY THE COMPONENTS OF A PROSTHESIS MIGRATE INTO THE SYNOVIAL CAVITY AND THE BONE-IMPLANT INTERFACE, WHERE THEY RECRUIT MACROPHAGES AND STIMULATE PHAGOCYTIC ACTION. THE DEGREE OF RECRUITMENT IS DETERMINED BY THE SIZE, DISTRIBUTION AND AMOUNT OF PARTICULATE DEBRIS (RATE OF DEBRIS GENERATION). THE PHAGOCYTIC ACTION RESULTS IN THE RELEASE OF CYTOKINES AND INTERCELLULAR MEDIATORS (IL-1, 2, PE2) WHICH ENCOURAGE OSTEOCLASTIC BONE RESORPTION. CLINICAL AND BASIC RESEARCH IS CONTINUING IN ORDER TO PROVIDE SCIENTIFIC BASIS FOR THE CAUSES OF THIS PHENOMENON AND THE POTENTIAL WAYS TO REDUCE ITS OCCURRENCE. OSTEOLYSIS CAN BE ASYMPTOMATIC AND THEREFORE ROUTINE PERIODIC RADIOGRAPHIC EXAMINATION IS VITAL TO PREVENT ANY SERIOUS FUTURE COMPLICATION. PRESENCE OF FOCAL LESIONS THAT ARE PROGRESSIVE MAY NECESSITATE REPLACEMENT OF THE PROSTHETIC COMPONENT(S).

MR SAFETY INFORMATION

THE APEX 3D™ TOTAL ANKLE REPLACEMENT SYSTEM HAS NOT BEEN EVALUATED FOR SAFETY AND COMPATIBILITY IN THE MR ENVIRONMENT. IT HAS NOT BEEN TESTED FOR HEATING, MIGRATION, OR IMAGE ARTIFACT IN THE MR ENVIRONMENT. THE SAFETY OF THE APEX 3D™ TOTAL ANKLE REPLACEMENT SYSTEM IN THE MR ENVIRONMENT IS UNKNOWN. MR SCANNING OF A PATIENT WHO HAS THIS DEVICE MAY RESULT IN PATIENT INJURY.




P10-STG-0010 Rev A [2022-05-09]

™Trademarks and ®Registered Marks of Paragon 28®, Inc.

© Copyright 2022 Paragon 28®, Inc. All rights reserved.

Patents: www.paragon28.com/patents

Paragon 28, Inc. 

14445 Grasslands Dr.

Englewood, CO 80112

(855) 786-2828

**PATENTED, DESIGNED & EXCLUSIVELY
DISTRIBUTED BY**

Exclusively foot & ankle ²⁸
Paragon®

www.Paragon28.com

DISCLAIMER

The purpose of the APEX 3D™ Total Ankle Replacement System Surgical Technique Guide is to demonstrate the use of the APEX 3D™ Total Ankle Replacement System. Although various methods can be employed for this procedure, the fixation options demonstrated were chosen for simplicity of explanation and demonstration of the unique features of our device. Federal law (U.S.A.) restricts this device to sale and use by, or on order of, a physician.